



## The GlideScope® 4-Step Technique



Looking directly into the patient's mouth and with the GlideScope® in the left hand, introduce the video laryngoscope into the midline of the oral pharynx.



With the laryngoscope inserted, look to the monitor to identify the epiglottis, then manipulate the scope to obtain the best glottic view.



Looking directly into the patient's mouth, not at the screen, carefully guide the distal tip of the tube into position near the tip of the laryngoscope.

*It is important to look into the mouth at this step to avoid injuring the tonsils or soft palate.*



Look to the monitor to complete the intubation; gently rotate or angle the tube to redirect as needed.

- Verathon® recommends inserting the GlideScope® Video Laryngoscope down the midline of the tongue to the epiglottis.
- The GlideScope® video laryngoscope may be used to produce a MacIntosh indirect lift of the epiglottis or a Miller lift.
- Intubations using GlideScope® Video Laryngoscopes require approximately 0.5kg-1.5kg of lifting force.


- Use of an endotracheal tube stylet is recommended. The GlideRite® Rigid Stylet has been designed to complement the angle of the GlideScope® video laryngoscope to facilitate intubation. A malleable stylet may be used with a 60°-90° angle.



- To aid the passage of the endotracheal tube, withdraw the stylet (approx. 5 cm) while gently advancing the ETT. A 1 cm adjustment (withdrawal) of the laryngoscope also may be beneficial to reduce the viewing angle and allow the glottis to drop.

Corporate Headquarters:  
Verathon Inc.  
20001 North Creek Parkway  
Bothell, WA 98011, USA  
Tel: 800.331.2313 (US and Canada only)  
Tel: 425.867.1348 Fax: 425.883.2896

 Verathon Medical  
(Canada) ULC  
4224 Manor Street  
Burnaby, British Columbia  
Canada, V5G 1B2

 Verathon Medical (Europe) B.V.  
Linnaeusweg 11  
3401 MS IJsselstein Netherlands  
Tel: +31.30.68.70.570  
Fax: +31.30.68.70.512t

 VERATHON  
MEDICAL  
verathon.com